



WECK LABORATORIES, INC.

Certificate of Analysis

FINAL REPORT

Work Orders: 8I05014

Report Date: 9/17/2018

Project: City of Paramount Cr6

Received Date: 9/7/2018

Turnaround Time: 4 workdays

Phones: (562) 275-4252

Fax: (562) 921-6101

Attn: Charlene King

P.O. #:

Client: Water Replenishment District
4040 Paramount Blvd.
Lakewood, CA 90712

Billing Code:

DoD-ELAP #L2457 • ELAP-CA #1132 • EPA-UCMR #CA00211 • Guam-EPA #17-008R • HW-DOH # • ISO 17025 #L2457.01 • LACSD #10143 • NELAP-CA #04229CA • NELAP-OR #4047 • NJ-DEP #CA015 • NV-DEP #NAC 445A • SCAQMD #93LA1006

This is a complete final report. The information in this report applies to the samples analyzed in accordance with the chain-of-custody document. Weck Laboratories certifies that the test results meet all requirements of TNI unless noted by qualifiers or written in the Case Narrative. This analytical report must be reproduced in its entirety.


Dear Charlene King,

Enclosed are the results of analyses for samples received 9/07/18 with the Chain-of-Custody document. The samples were received in good condition, at 4.2 °C and on ice. All analyses met the method criteria except as noted in the case narrative or in the report with data qualifiers.

Reviewed by:

Valerie Rejuso
Project Manager





WECK LABORATORIES, INC.

Water Replenishment District
4040 Paramount Blvd.
Lakewood, CA 90712

Certificate of Analysis

FINAL REPORT

Project Number: City of Paramount Cr6


Reported:

09/17/2018 14:42

Project Manager: Charlene King

Sample Summary

Sample Name	Sampled By	Lab ID	Matrix	Sampled	Qualifiers
1910105020/EPTDS from MWD	Carlos Navarro	8I05014-01	Water	09/07/18 10:58	
1910105023W14D/Well 14 Disinfection Station	Carlos Navarro	8I05014-03	Water	09/07/18 10:05	
1910105994/DSMRT for Well 15	Carlos Navarro	8I05014-04	Water	09/07/18 10:49	
1910105024W15D/Well 15 Disinfection Station	Carlos Navarro	8I05014-05	Water	09/07/18 11:10	
1910105991/DSMRT for MWD	Carlos Navarro	8I05014-06	Water	09/07/18 10:40	
1910105992/DSMRT for Well 13	Carlos Navarro	8I05014-07	Water	09/07/18 10:25	
19105993/DSMRT for Well 14	Carlos Navarro	8I05014-08	Water	09/07/18 10:14	



WECK LABORATORIES, INC.

Water Replenishment District
4040 Paramount Blvd.
Lakewood, CA 90712

Certificate of Analysis

FINAL REPORT

Project Number: City of Paramount Cr6

Reported:

09/17/2018 14:42

Project Manager: Charlene King

Sample Results

Sample: 1910105020/EPTDS from MWD
8I05014-01 (Water) Sampled: 09/07/18 10:58 by Carlos Navarro

Analyte	Result	MRL	Units	Dil	Analyzed	Qualifier
Hexavalent Chromium by IC						
Method: EPA 218.6	Batch ID: W8I0653	Instr: LC13	Prepared: 09/13/18 09:36	Analyst: blg		
Chromium 6+	0.046	0.020	ug/l	1	09/14/18 18:02	

Sample: 1910105023W14D/Well 14 Disinfection Station
8I05014-03 (Water) Sampled: 09/07/18 10:05 by Carlos Navarro

Analyte	Result	MRL	Units	Dil	Analyzed	Qualifier
Hexavalent Chromium by IC						
Method: EPA 218.6	Batch ID: W8I0653	Instr: LC13	Prepared: 09/13/18 09:36	Analyst: blg		
Chromium 6+	0.17	0.020	ug/l	1	09/14/18 18:14	

Sample: 1910105994/DSMRT for Well 15
8I05014-04 (Water) Sampled: 09/07/18 10:49 by Carlos Navarro

Analyte	Result	MRL	Units	Dil	Analyzed	Qualifier
Hexavalent Chromium by IC						
Method: EPA 218.6	Batch ID: W8I0653	Instr: LC13	Prepared: 09/13/18 09:36	Analyst: blg		
Chromium 6+	ND	0.020	ug/l	1	09/14/18 18:25	

Sample: 1910105024W15D/Well 15 Disinfection Station
8I05014-05 (Water) Sampled: 09/07/18 11:10 by Carlos Navarro


Analyte	Result	MRL	Units	Dil	Analyzed	Qualifier
Hexavalent Chromium by IC						
Method: EPA 218.6	Batch ID: W8I0653	Instr: LC13	Prepared: 09/13/18 09:36	Analyst: blg		
Chromium 6+	ND	0.020	ug/l	1	09/14/18 18:37	

Sample: 1910105991/DSMRT for MWD
8I05014-06 (Water) Sampled: 09/07/18 10:40 by Carlos Navarro

Analyte	Result	MRL	Units	Dil	Analyzed	Qualifier
Hexavalent Chromium by IC						
Method: EPA 218.6	Batch ID: W8I0653	Instr: LC13	Prepared: 09/13/18 09:36	Analyst: blg		
Chromium 6+	0.15	0.020	ug/l	1	09/14/18 18:49	

Sample: 1910105992/DSMRT for Well 13
8I05014-07 (Water) Sampled: 09/07/18 10:25 by Carlos Navarro

Analyte	Result	MRL	Units	Dil	Analyzed	Qualifier
Hexavalent Chromium by IC						
Method: EPA 218.6	Batch ID: W8I0653	Instr: LC13	Prepared: 09/13/18 09:36	Analyst: blg		
Chromium 6+	0.075	0.020	ug/l	1	09/14/18 19:01	



WECK LABORATORIES, INC.

Water Replenishment District
4040 Paramount Blvd.
Lakewood, CA 90712

Certificate of Analysis

FINAL REPORT

Project Number: City of Paramount Cr6

Reported:
09/17/2018 14:42

Project Manager: Charlene King




Sample Results

(Continued)

Sample: 19105993/DSMRT for Well 14
8I05014-08 (Water)

Sampled: 09/07/18 10:14 by Carlos Navarro

Analyte	Result	MRL	Units	Dil	Analyzed	Qualifier
Hexavalent Chromium by IC						
Method: EPA 218.6	Batch ID: W8I0653	Instr: LC13	Prepared: 09/13/18 09:36	Analyst: blg		
Chromium 6+	0.18	0.020	ug/l	1	09/14/18 19:13	



WECK LABORATORIES, INC.

Water Replenishment District
4040 Paramount Blvd.
Lakewood, CA 90712

Certificate of Analysis

FINAL REPORT

Project Number: City of Paramount Cr6

Reported:


09/17/2018 14:42

Project Manager: Charlene King

Quality Control Results

Hexavalent Chromium by IC

Analyte	Result	MRL	Units	Spike Level	Source Result	%REC	Limits	RPD	RPD Limit	Qualifier
Batch: W810653 - EPA 218.6										
Blank (W810653-BLK1)										
Chromium 6+	ND	0.020	ug/l							
				Prepared: 09/13/18 Analyzed: 09/14/18						
LCS (W810653-BS1)										
Chromium 6+	5.07	0.020	ug/l	5.00		101	90-110			
				Prepared: 09/13/18 Analyzed: 09/14/18						
Matrix Spike (W810653-MS1)										
Chromium 6+	4.98	0.020	ug/l	5.00	0.0463	99	88-112			
				Prepared: 09/13/18 Analyzed: 09/14/18						
Matrix Spike Dup (W810653-MSD1)										
Chromium 6+	5.16	0.020	ug/l	5.00	0.0463	102	88-112	4	10	
				Prepared: 09/13/18 Analyzed: 09/14/18						



WECK LABORATORIES, INC.

Water Replenishment District
4040 Paramount Blvd.
Lakewood, CA 90712

Project Number: City of Paramount Cr6

Project Manager: Charlene King

Certificate of Analysis

FINAL REPORT

Reported:

09/17/2018 14:42



Notes and Definitions

Item	Definition
ND	NOT DETECTED at or above the Method Reporting Limit (MRL). If Method Detection Limit (MDL) is reported, then ND means not detected at or above the MDL.
Dil	Dilution
dry	Sample results reported on a dry weight basis
RPD	Relative Percent Difference
% Rec	Percent Recovery
Source	Sample that was matrix spiked or duplicated.
MDL	Method Detection Limit
MRL	The minimum levels, concentrations, or quantities of a target variable (e.g., target analyte) that can be reported with a specified degree of confidence. The MRL is also known as Limit of Quantitation (LOQ) and Detection Limit for Reporting (DLR)
MDA	Minimum Detectable Activity
NR	Not Reportable
TIC	Tentatively Identified Compound (TIC) using mass spectrometry. The reported concentration is relative concentration based on the nearest internal standard. If the library search produces no matches at, or above 85%, the compound is reported as unknown.

Any remaining sample(s) will be disposed of one month from the final report date unless other arrangements are made in advance.

An Absence of Total Coliform meets the drinking water standards as established by the California State Water Resources Control Board (SWRCB)

All results are expressed on wet weight basis unless otherwise specified.

All samples collected by Weck Laboratories have been sampled in accordance to laboratory SOP Number MIS 002.



Weck Laboratories, Inc.
Analytical Laboratory Services - Since 1964

CHAIN OF CUSTODY RECORD

14859 East Clark Avenue : Industry : CA 91745
Tel 626-336-2139 ♦ Fax 626-336-2634 ♦ www.wecklabs.com

STANDARD **8105014**

Page 1 Of 1

CLIENT NAME: Water Replenishment District		PROJECT: City of Paramount		ANALYSES REQUESTED				SPECIAL HANDLING	
ADDRESS: 4040 Paramount Blvd Lakewood, Ca 90712		PHONE: 562-921-5521 FAX: EMAIL: cking@wrd.org						<input type="checkbox"/> Same Day Rush 150% <input type="checkbox"/> 24 Hour Rush 100% <input type="checkbox"/> 48-72 Hour Rush 75% <input checked="" type="checkbox"/> 4 - 5 Day Rush 30% <input type="checkbox"/> Rush Extractions 50% <input type="checkbox"/> 10 - 15 Business Days <input type="checkbox"/> QA/QC Data Package	
PROJECT MANAGER: Charlene King		SAMPLER: Weck <i>Carlson</i>						Charges will apply for weekends/holidays	

ID# (For Lab Use Only)	DATE SAMPLED	TIME SAMPLED	SMPL TYPE	SAMPLE IDENTIFICATION/SITE LOCATION	# OF CONT.	C16 218.6										COMMENTS		
	9/7/18	1058	GW	1910105020/EPTDS from MWD	1	x												Fld Temp 28.2c
			GW	1910105022W13D/Well 13 Disinfection Station	1	x												Fld Temp not operating
		1005	GW	1910105023W14D/Well 14 Disinfection Station	1	x												Fld Temp 22.2c
		1049	GW	1910105994/DSMRT for Well 15 1910105024W15D/Well 15 Disinfection Station	1	x												Fld Temp 26.1c
		1110	GW	1910105991/DSMRT for MWD	1	x												Fld Temp 24.1c
		1040	GW	1910105992/DSMRT for Well 13	1	x												Fld Temp 26.2c
		1025	GW	1910105993/DSMRT for Well 14	1	x												Fld Temp 26.9c
		1014	GW		1	x												Fld Temp 26.1c

RELINQUISHED BY <i>Carlson</i>	DATE / TIME 9/7/18 1205	RECEIVED BY <i>Janakumar</i>	9/7/18 1205	SAMPLE CONDITION: Actual Temperature: 4.2°C <input checked="" type="checkbox"/> Received On Ice <input checked="" type="checkbox"/> Preserved <input checked="" type="checkbox"/> Evidence Seals Present <input checked="" type="checkbox"/> Container Attacked <input checked="" type="checkbox"/> Preserved at Lab	SAMPLE TYPE CODE: AQ=Aqueous NA= Non Aqueous SL = Sludge DW = Drinking Water WW = Waste Water RW = Rain Water GW = Ground Water SO = Soil SW = Solid Waste OL = Oil OT = Other Matrix
RELINQUISHED BY	DATE / TIME	RECEIVED BY			
RELINQUISHED BY	DATE / TIME	RECEIVED BY			

PRESCHEDULED RUSH ANALYSES WILL TAKE PRIORITY OVER UNSCHEDULED RUSH REQUESTS

Client agrees to Terms & Conditions at: www.wecklabs.com

SPECIAL REQUIREMENTS / BILLING INFORMATION

COC version 04/27/17